Table	41	(continue
IAUIC	71	Continue

	· · · · · · · · · · · · · · · · · · ·	lable 4(continued)							
Signaling Indicator response threshold correvalue, sponding			S4 positions						
		S3 positions	S4.1	S4.2	S4.3	S4.4			
μSv/h	to the		800	400	200				
(μ R/h)	threshold		800	400	200	100			
4 (400)	0400	Upper	1	1	0	0			
5 (500)	0500.	-11-	1	0	ı	l i			
6 (600)	0600	-11-	1	Ŏ	l i	o.			
7 (700)	0700	-11-	1	0	0	i			
8 (000)	0800	- ii -	1	0	0	Ò			
9 (900)	0900	-ii-	0	1	1	i			
10 (1000)	1000	- ii -	0	1	1	ō			
11 (1100)	1100	-11-	0	1	0	1			
12 (1200)	1200	- ii i	0	1	0	0			
13 (1300)	1300	- ii -	0	0	ì	i			
14 (1400)	1400	- jj -	0	0	1	ō			
15 (1500)	1500	-11-	0	0	0	i			
16 (1600)	1600	-11-1	0	0	. 0	0			

8. SERVICE INSTRUCTIONS. GENERAL

- 8.1. Keep the instrument clean, regularly dust it with dry and clean cloth.
- 8.2. Protect the instrument from shocks and mechanical damage.
- 8.3. The instrument is equipped with a removable filter-cover 4 which is removed when taking radiometric measurements. To avoid damage remove and replace the cover carefully in compliance with Section 7 of the present manual.

8.4. To avoid contamination of the casing with

- radionuclides and to protect it from ingress of moisture it is advisable to keep the instrument inside a plastic cover.

 8.5. To avoid waste of power supply de-energise
- the instrument during the idle period.

 8.6. The manufacturer does not supply spare power supply sources, they have to be purchased by user on his own. To replace the power supply source
- a) remove cover of power supply compartment by shifting it in the direction shown by the arrow; b) extract the power block with connected battery

proceed as follows:

- b) extract the power block with connected battery and, holding the block, disconnect the battery;
 c) connect a fresh battery to the power block as it
- is specified in Section 6; replace the power block with fresh battery and replace the cover.

 8.7. As a result of prolonged operation of the instrument In areas with excessive radio-contamination the instrument can override even if
- sources of ionizing radiation are not present in the vicinity. In that case the instrument should be decontaminated in the following order:

 a) prepare neutral decontamination solution
- a) prepare neutral decontamination solution (dissolve 1/3 of tea-spoon of neutral detergent in 0.5 l of lukewarm clean water);